

INDIVIDUAL LABORATORY REPORT (ILR) - Lab 1792

SCHEME: Oncogene panel testing

SEASON: 2016

Case 1		
Assessment Category	Score ¹	Comments (& deductions ²)
Genotyping	2.00	No comments (0)

Case 2		
Assessment Category	Score ¹	Comments (& deductions ²)
Genotyping	2.00	No comments (0)

Case 3		
Assessment Category	Score ¹	Comments (& deductions ²)
Genotyping	2.00	No comments (0) Missed EGFR 2369C>T (1.02%) ? below LOD. Missed PIK3CA c.1633G>A (7.8%) - above LOD but exon 9 not tested.

Case 4		
Assessment Category	Score ¹	Comments (& deductions ²)
Genotyping	--	Critical genotyping error Missed EGFR 2235_2249del (2.3%) and c.2369C>T 0.85%). The deletion should have been detected at 2% LOD False negative result Using NGS

General Comments	Thank you for participating in this years EQA scheme. We look forward to you joining us in the 2017 scheme - registration to participate will be available during January 2017 via the EMQN website
-------------------------	---

SUMMARY OF YOUR PERFORMANCE IN THIS SCHEME

Assessment Category	Performance ³ (mean score)
Genotyping	2.00
Scheme result (SATISFACTORY or POOR)	SATISFACTORY

NOTE: The results above are subject to change depending on the results of the appeals process. The final results will be displayed on the EMQN certificate of participation

Report approved and authorised by Simon Patton (10 November 2016) on behalf of EMQN.
 Signed:



This report is provisional pending the outcome of the appeals process

¹ Maximum score is 2.00

² Deductions from the maximum score

³ Green >= Scheme mean, Orange < Scheme mean, Red Poor performance.

NRS no results submitted. WDS withdrew from scheme